




Defibrillator Compatibility

Match Defibrillator Brands to Vermed Part Numbers

BRAND OF DEFIBRILLATOR	MODEL OF DEFIBRILLATOR	VERMED COMPATIBLE DEFIBRILLATION ELECTRODE PART NUMBERS
Philips Medical	• XLT • HeartStart XL • HP Codemaster XL (Plug Style Connector)	• T100AC-Philips • T100LOAC-Philips • C100AC-Philips
	• HeartStart FR2+ • HeartStart Forerunner	• T100AC-Philips • T100LOAC-Philips • C100AC-Philips • C100LOAC-Philips
	• HeartStart FR3 • HeartStart FRx	NOT COMPATIBLE*
Physio-Control	• LP-9P • LP-10C • LP-12	• T100AC-Physio • T100LOAC-Physio • C100AC-Physio • C100LOAC-Physio • Pediatric-Physio
	• LP-500 • LP-1000	• T100AC-Physio • T100LOAC-Physio • C100AC-Physio • C100LOAC-Physio
Zoll Medical	• M-Series • R-Series	• T100AC-Zoll • T100LOAC-Zoll • C100AC-Zoll • C100LOAC-Zoll • Pediatric-Zoll
	• Zoll AED PLUS • Zoll AED Pro	NOT COMPATIBLE*
Cardiac Science	• PowerHeart AED G3 Plus • PowerHeart AED G3 Pro	• T100-CS

*HeartSync pediatric defibrillation electrodes are NOT compatible with AED's. AED's cannot be manually adjusted to reduce energy delivery. Please refer to your AED's manufacturer operation manual.

Please contact Vermed directly for any questions regarding electrode compatibility:

400 Exchange Street | Buffalo, NY 14204 | Phone: 800.669.6905 | Fax: 888.658.4941 | vermed.com